

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 007256  
10:37 24/04/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:32  
PRESSURIZING 2:33  
STERILIZING 3:32  
136.4 C 220kPa 3:32  
136.3 C 223kPa 4:02  
136.4 C 224kPa 4:32  
136.4 C 223kPa 5:02  
136.4 C 223kPa 5:32  
136.4 C 222kPa 6:02  
136.2 C 223kPa 6:32  
136.3 C 224kPa 7:02  
136.3 C 224kPa 7:02  
VENTING 7:02  
AIR DRYING 7:53  
CYCLE COMPLETE 37:53  
-----

Digital Signature #  
6E718F9D6D669EA5  
-----