
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002066
10:23 19/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:39
PRESSURIZING 3:40
STERILIZING 4:50
136.4 C 220kPa 4:50
136.2 C 222kPa 5:20
136.4 C 223kPa 5:50
136.3 C 223kPa 6:20
136.1 C 220kPa 6:50
135.9 C 220kPa 7:20
136.4 C 224kPa 7:50
136.2 C 222kPa 8:20
136.2 C 222kPa 8:20
VENTING 8:20
AIR DRYING 9:16
CYCLE COMPLETE 9:17

Digital Signature #
2C77556BE1661F18
