
Midmark M11 - v1.0.5
Total Cycles: 3095

Name: RITTER2
Sterilizer ID: V2313228
11/07/2023 03:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.8 F 0.0
2:00 120.0 F 0.6
4:00 133.6 F 3.0
6:00 166.1 F 6.5
8:00 189.2 F 10.2
10:00 205.7 F 7.2
12:00 221.8 F 7.5
14:00 230.8 F 12.2
16:00 234.1 F 9.2
18:00 238.6 F 10.4
20:00 246.0 F 13.9
22:00 257.9 F 20.3
24:00 268.6 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.8 F 29.9
2:00 271.5 F 29.2
3:00 272.0 F 30.0
4:00 272.5 F 29.6

Min 271.1 F 29.1
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.5 F 0.0
5:00 204.2 F 0.0
10:00 198.5 F 0.0
15:00 196.2 F 0.0
20:00 195.2 F 0.0
25:00 193.9 F 0.0
30:00 192.9 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 25:12
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:03:46
CYCLE COMPLETE

11/07/2023 04:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____