

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2039

Name: RITTER 1  
Sterilizer ID: V2178198  
06/21/2022 08:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 119.7 F 0.0  
2:00 124.9 F 0.0  
4:00 153.3 F 1.3  
6:00 166.7 F 3.9  
8:00 186.3 F 7.7  
10:00 203.8 F 11.8  
12:00 216.1 F 4.5  
14:00 228.5 F 8.8  
16:00 237.9 F 13.3  
18:00 237.2 F 10.2  
20:00 236.2 F 8.2  
22:00 247.6 F 14.2  
24:00 259.2 F 20.5  
26:00 269.6 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.5 F 29.2  
2:00 271.8 F 28.4  
3:00 271.9 F 29.0  
4:00 272.4 F 29.0

Min 271.4 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.3 F 0.0  
5:00 198.1 F 0.0  
10:00 209.3 F 0.0  
15:00 212.2 F 0.0  
20:00 215.4 F 0.0  
25:00 214.5 F 0.0  
30:00 216.9 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 27:03  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:05:07  
CYCLE COMPLETE

06/21/2022 10:01 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_