
Midmark M11 - v1.0.5
Total Cycles: 1366

Name: RITTER2
Sterilizer ID: V2313228
08/04/2022 06:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.9 F 0.0
2:00 75.3 F 0.0
4:00 84.6 F 0.0
6:00 99.3 F 0.1
8:00 131.2 F 2.0
10:00 160.6 F 5.4
12:00 185.2 F 9.0
14:00 207.1 F 12.7
16:00 214.8 F 6.3
18:00 226.1 F 10.3
20:00 231.8 F 8.9
22:00 237.6 F 11.3
24:00 235.8 F 9.2
26:00 248.2 F 15.2
28:00 259.0 F 21.2
30:00 268.8 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.4
1:00 272.7 F 30.1
2:00 271.9 F 29.3
3:00 272.1 F 30.1
4:00 272.5 F 29.8

Min 271.4 F 29.2
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.9 F 0.0
5:00 200.6 F 0.0
10:00 210.1 F 0.0
15:00 212.9 F 0.0
20:00 223.9 F 0.0
25:00 229.4 F 0.0
30:00 236.5 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 31:14
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:09:17
CYCLE COMPLETE

08/04/2022 07:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____