
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 005190
10:03 02/03/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:05
PRESSURIZING 2:09
STERILIZING 3:17
136.4 C 221kPa 3:17
136.3 C 222kPa 3:47
136.4 C 224kPa 4:17
136.3 C 220kPa 4:47
136.2 C 220kPa 5:17
136.2 C 223kPa 5:47
136.3 C 220kPa 6:17
136.4 C 223kPa 6:47
136.4 C 223kPa 6:47
VENTING 6:47
AIR DRYING 7:36
CYCLE COMPLETE 10:46

Digital Signature #
F0DCC797B7F09F03
