
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001475
12:21 25/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:32
STERILIZING 3:40
136.4 C 224kPa 3:40
136.2 C 225kPa 4:10
136.4 C 226kPa 4:40
136.3 C 225kPa 5:10
136.3 C 224kPa 5:40
136.3 C 226kPa 6:10
136.4 C 225kPa 6:40
136.2 C 225kPa 7:10
136.2 C 225kPa 7:10
VENTING 7:10
AIR DRYING 8:03
CYCLE COMPLETE 8:28

Digital Signature #
BCA3DEB8E8CA1777
