
Midmark M11 - v1.0.5
Total Cycles: 1427

Name: RITTER2
Sterilizer ID: V2313228
08/29/2022 01:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.7 F 0.0
2:00 184.7 F 2.9
4:00 193.7 F 7.7
6:00 220.1 F 13.4
8:00 227.4 F 9.1
10:00 234.1 F 9.6
12:00 243.9 F 13.4
14:00 244.5 F 12.9
16:00 257.9 F 20.1
18:00 269.4 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 271.9 F 29.1
2:00 272.3 F 29.5
3:00 271.8 F 29.6
4:00 271.6 F 29.1

Min 271.3 F 29.0
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.7 F 0.0
5:00 218.8 F 0.0
10:00 239.1 F 0.0
15:00 237.6 F 0.0
20:00 227.9 F 0.0
25:00 220.5 F 0.0
30:00 218.6 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 18:54
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:41
CYCLE COMPLETE

08/29/2022 02:46 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____