

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2328

Name: RITTER 1  
Sterilizer ID: V2178198  
09/15/2022 07:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 76.6 F 0.0  
2:00 79.0 F 0.0  
4:00 86.2 F 0.0  
6:00 108.7 F 0.0  
8:00 145.2 F 0.1  
10:00 175.8 F 2.7  
12:00 192.9 F 6.0  
14:00 205.3 F 9.2  
16:00 217.6 F 12.7  
18:00 228.0 F 6.4  
20:00 238.1 F 10.8  
22:00 234.7 F 7.5  
24:00 245.9 F 13.2  
26:00 243.6 F 11.7  
28:00 256.0 F 18.1  
30:00 266.9 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 27.9  
1:00 272.5 F 28.6  
2:00 271.3 F 27.8  
3:00 272.4 F 28.8  
4:00 271.8 F 28.1  
  
Min 271.3 F 27.8  
Max 272.6 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.7 F 0.0  
5:00 203.1 F 0.0  
10:00 227.2 F 0.0  
15:00 225.7 F 0.0  
20:00 217.5 F 0.0  
25:00 218.4 F 0.0  
30:00 214.7 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 31:34  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:09:34  
CYCLE COMPLETE

09/15/2022 08:10 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_