
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000072
11:48 31/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:36
PRESSURIZING 3:39
STERILIZING 4:56
136.4 C 224kPa 4:56
136.2 C 224kPa 5:26
136.2 C 223kPa 5:56
136.4 C 224kPa 6:26
136.3 C 224kPa 6:56
136.3 C 225kPa 7:26
136.1 C 222kPa 7:56
136.4 C 224kPa 8:26
136.4 C 224kPa 8:26
VENTING 8:26
AIR DRYING 9:19
CYCLE COMPLETE 24:19

Digital Signature #
52941DB32EB5BFF2
