

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002351  
10:59 04/12/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:55  
PRESSURIZING 3:00  
STERILIZING 4:05  
136.4 C 221kPa 4:05  
136.2 C 223kPa 4:35  
136.4 C 224kPa 5:05  
136.2 C 222kPa 5:35  
136.3 C 223kPa 6:05  
136.2 C 223kPa 6:35  
136.2 C 222kPa 7:05  
136.0 C 221kPa 7:35  
136.0 C 221kPa 7:35  
VENTING 7:35  
AIR DRYING 8:24  
CYCLE COMPLETE 38:24  
-----

Digital Signature #  
8CF23F7E131CEDBD  
-----