
Midmark M11 - v1.0.5
Total Cycles: 3137

Name: RITTER2
Sterilizer ID: V2313228
11/17/2023 10:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.2 F 0.0
2:00 69.3 F 0.0
4:00 71.0 F 0.2
6:00 86.8 F 1.3
8:00 127.2 F 3.6
10:00 152.3 F 6.6
12:00 178.2 F 10.0
14:00 196.8 F 13.4
16:00 207.3 F 5.7
18:00 219.0 F 9.4
20:00 230.9 F 13.5
22:00 233.9 F 10.1
24:00 236.5 F 9.4
26:00 245.8 F 13.9
28:00 256.7 F 19.8
30:00 266.7 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.5
1:00 272.6 F 30.2
2:00 271.2 F 29.3
3:00 272.8 F 30.1
4:00 271.1 F 29.4

Min 271.1 F 29.3
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.6 F 0.0
5:00 189.1 F 0.0
10:00 184.7 F 0.0
15:00 179.1 F 0.0
20:00 175.8 F 0.0
25:00 175.6 F 0.0
30:00 174.9 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 31:27
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:10:02
CYCLE COMPLETE

11/17/2023 11:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____