
Midmark M11 - v1.0.5
Total Cycles: 3099

Name: RITTER2
Sterilizer ID: V2313228
11/08/2023 11:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.0 F 0.0
2:00 178.9 F 2.4
4:00 191.9 F 7.1
6:00 209.5 F 11.9
8:00 224.0 F 7.1
10:00 238.0 F 12.4
12:00 237.3 F 10.3
14:00 237.1 F 9.0
16:00 250.7 F 16.0
18:00 262.3 F 22.7
20:00 272.8 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.7 F 29.7
2:00 272.0 F 29.1
3:00 271.1 F 29.0
4:00 272.3 F 29.7

Min 271.0 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.4 F 0.0
5:00 234.2 F 0.0
10:00 202.5 F 0.0
15:00 196.8 F 0.0
20:00 194.9 F 0.0
25:00 195.5 F 0.0
30:00 199.1 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 20:24
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 00:59:36
CYCLE COMPLETE

11/08/2023 12:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____