

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1042

Name: RITTER 1  
Sterilizer ID: V2178198  
07/22/2021 06:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.5 F 0.0  
2:00 74.0 F 0.0  
4:00 82.5 F 0.0  
6:00 130.1 F 0.1  
8:00 158.3 F 0.1  
10:00 187.5 F 0.2  
12:00 202.5 F 1.0  
14:00 221.6 F 4.2  
16:00 228.3 F 6.7  
18:00 235.8 F 9.8  
20:00 243.4 F 12.4  
22:00 239.5 F 10.4  
24:00 236.2 F 8.2  
26:00 244.5 F 11.7  
28:00 246.0 F 12.9  
30:00 258.0 F 19.3  
32:00 268.6 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.0  
1:00 272.6 F 28.7  
2:00 271.5 F 27.8  
3:00 272.4 F 28.8  
4:00 272.0 F 28.3  
  
Min 271.2 F 27.8  
Max 272.6 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.4 F 0.0  
5:00 213.6 F 0.0  
10:00 210.9 F 0.0  
15:00 227.1 F 0.0  
20:00 219.5 F 0.0  
25:00 217.1 F 0.0  
30:00 179.8 F 0.0  
  
Drying Complete

FILLING: 1:08  
HEATING: 33:14  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:11:28  
CYCLE COMPLETE

07/22/2021 07:57 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02