

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3803

Name: RITTER2  
Sterilizer ID: V2313228  
07/22/2024 08:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.7 F 0.0  
2:00 78.2 F 0.0  
4:00 90.6 F 0.0  
6:00 110.9 F 0.1  
8:00 134.3 F 0.5  
10:00 164.2 F 3.2  
12:00 197.5 F 6.4  
14:00 207.6 F 10.4  
16:00 223.3 F 7.2  
18:00 228.5 F 7.7  
20:00 234.9 F 11.9  
22:00 234.0 F 8.5  
24:00 243.7 F 12.2  
26:00 243.9 F 12.4  
28:00 255.6 F 18.5  
30:00 265.9 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 272.4 F 29.2  
2:00 271.6 F 29.0  
3:00 272.0 F 28.9  
4:00 271.9 F 29.3

Min 271.3 F 28.5  
Max 272.6 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.5 F 0.0  
5:00 228.6 F 0.0  
10:00 230.0 F 0.0  
15:00 229.7 F 0.0  
20:00 229.5 F 0.0  
25:00 223.8 F 0.0  
30:00 228.5 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 31:45  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:10:05  
CYCLE COMPLETE

07/22/2024 10:08 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_