
Midmark M11 - v1.0.5
Total Cycles: 1888

Name: RITTER2
Sterilizer ID: V2313228
01/03/2023 08:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.7 F 0.0
2:00 79.0 F 0.0
4:00 89.2 F 0.0
6:00 118.4 F 0.1
8:00 164.8 F 2.0
10:00 198.6 F 6.1
12:00 212.9 F 10.8
14:00 222.5 F 5.5
16:00 231.9 F 9.9
18:00 238.4 F 9.1
20:00 243.3 F 12.6
22:00 241.9 F 11.3
24:00 255.4 F 18.3
26:00 267.1 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 271.7 F 28.7
2:00 271.9 F 28.9
3:00 272.0 F 29.0
4:00 272.1 F 29.1

Min 271.6 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 232.6 F 0.0
Dry Cycle Aborted

FILLING: 0:44
HEATING: 27:16
STERILIZING: 4:00
VENTING: 2:39
DRYING: 2:01
TOTAL CYCLE: 00:36:40

01/03/2023 09:19 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____