
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007301
9:45 02/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:35
PRESSURIZING 2:37
STERILIZING 3:32
136.4 C 222kPa 3:32
136.4 C 223kPa 4:02
136.4 C 220kPa 4:32
136.3 C 222kPa 5:02
136.4 C 224kPa 5:32
136.4 C 222kPa 6:02
136.4 C 223kPa 6:32
136.1 C 217kPa 7:02
136.1 C 217kPa 7:02
VENTING 7:02
AIR DRYING 7:53
CYCLE COMPLETE 21:37

Digital Signature #
DCC68C15A579526E
