
Midmark M11 - v1.0.5
Total Cycles: 2830

Name: RITTER2
Sterilizer ID: V2313228
09/05/2023 11:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.9 F 0.0
2:00 181.6 F 3.3
4:00 194.7 F 8.0
6:00 210.9 F 12.9
8:00 226.9 F 8.2
10:00 241.4 F 12.9
12:00 240.5 F 11.7
14:00 239.9 F 10.8
16:00 253.1 F 17.3
18:00 264.1 F 23.8
20:00 271.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.4
1:00 272.4 F 29.5
2:00 272.2 F 29.8
3:00 271.6 F 29.7
4:00 272.0 F 29.5

Min 271.5 F 29.3
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.7 F 0.0
5:00 226.9 F 0.0
10:00 227.8 F 0.0
15:00 230.3 F 0.0
20:00 228.8 F 0.0
25:00 228.1 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 20:49
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:00:11
CYCLE COMPLETE

09/05/2023 12:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____