
Midmark M11 - v1.0.0
Total Cycles: 2504

Name: RITTER 1
Sterilizer ID: V2178198
11/02/2022 04:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.9 F 0.0
2:00 169.6 F 3.8
4:00 181.6 F 7.8
6:00 208.4 F 13.1
8:00 222.9 F 6.5
10:00 236.2 F 11.5
12:00 233.7 F 8.2
14:00 242.8 F 11.4
16:00 244.9 F 12.6
18:00 256.9 F 19.1
20:00 267.8 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.7 F 29.4
2:00 272.6 F 29.0
3:00 271.9 F 28.5
4:00 271.7 F 28.7

Min 271.5 F 28.3
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 233.0 F 0.0
10:00 229.2 F 0.0
15:00 230.1 F 0.0
20:00 226.0 F 0.0
25:00 229.1 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 21:25
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:38
CYCLE COMPLETE

11/02/2022 05:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____