
Midmark M11 - v1.0.0
Total Cycles: 1295

Name: RITTER 1
Sterilizer ID: V2178198
10/25/2021 08:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.2 F 0.0
2:00 106.6 F 0.0
4:00 124.7 F 0.1
6:00 145.8 F 0.1
8:00 179.0 F 0.4
10:00 207.6 F 3.7
12:00 212.7 F 6.9
14:00 223.7 F 10.6
16:00 231.0 F 7.6
18:00 237.0 F 10.6
20:00 236.4 F 8.3
22:00 246.4 F 13.0
24:00 244.4 F 12.3
26:00 256.3 F 18.6
28:00 267.2 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.6 F 28.8
2:00 271.6 F 28.3
3:00 272.6 F 28.7
4:00 271.5 F 28.1

Min 271.5 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.1 F 0.0
5:00 209.3 F 0.0
10:00 208.5 F 0.0
15:00 198.2 F 0.0
20:00 192.9 F 0.0
25:00 194.5 F 0.0
30:00 180.0 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 29:29
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:07:45
CYCLE COMPLETE

10/25/2021 10:05 AM
ACCEPTED: X
REJECTED:
Operator: U01