
Midmark M11 - v1.0.5
Total Cycles: 997

Name: RITTER2
Sterilizer ID: V2313228
04/18/2022 11:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.8 F 0.0
2:00 145.2 F 0.0
4:00 158.0 F 0.4
6:00 194.2 F 2.9
8:00 209.7 F 7.7
10:00 225.9 F 12.8
12:00 232.0 F 9.2
14:00 237.4 F 10.0
16:00 245.0 F 12.9
18:00 245.7 F 13.3
20:00 258.9 F 20.5
22:00 270.5 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.4 F 29.1
2:00 272.1 F 29.3
3:00 272.1 F 28.9
4:00 272.3 F 29.5

Min 271.5 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.8 F 0.0
5:00 195.3 F 0.0
10:00 197.7 F 0.0
15:00 194.8 F 0.0
20:00 191.3 F 0.0
25:00 191.8 F 0.0
30:00 204.2 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 22:46
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:54
CYCLE COMPLETE

04/18/2022 12:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____