

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3567

Name: RITTER 1  
Sterilizer ID: V2178198  
07/24/2023 09:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.3 F 0.0  
2:00 154.2 F 0.0  
4:00 164.7 F 2.1  
6:00 186.0 F 4.6  
8:00 202.9 F 8.1  
10:00 216.5 F 12.0  
12:00 219.9 F 5.4  
14:00 231.7 F 9.7  
16:00 237.3 F 10.8  
18:00 240.4 F 11.2  
20:00 237.2 F 9.2  
22:00 249.9 F 15.1  
24:00 260.4 F 21.0  
26:00 270.0 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.5 F 29.0  
2:00 272.1 F 28.5  
3:00 271.3 F 28.3  
4:00 272.4 F 29.0

Min 271.2 F 28.2  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.3 F 0.0  
5:00 202.8 F 0.0  
10:00 201.0 F 0.0  
15:00 199.0 F 0.0  
20:00 198.4 F 0.0  
25:00 195.6 F 0.0  
30:00 196.1 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 27:02  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:05:39  
CYCLE COMPLETE

07/24/2023 10:57 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_