
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001620
17:17 20/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 2:59
STERILIZING 3:59
136.4 C 224kPa 3:59
136.3 C 225kPa 4:29
136.3 C 226kPa 4:59
136.4 C 224kPa 5:29
136.4 C 227kPa 5:59
136.2 C 223kPa 6:29
136.4 C 226kPa 6:59
136.3 C 224kPa 7:29
136.3 C 224kPa 7:29
VENTING 7:29
AIR DRYING 8:25
CYCLE COMPLETE 8:32

Digital Signature #
2EA5A183DE8D8425
