
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007054
13:07 08/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:48
PRESSURIZING 2:52
STERILIZING 3:50
136.5 C 223kPa 3:50
136.3 C 221kPa 4:20
136.4 C 222kPa 4:50
136.4 C 225kPa 5:20
136.4 C 224kPa 5:50
136.2 C 223kPa 6:20
136.1 C 222kPa 6:50
136.4 C 225kPa 7:20
136.4 C 225kPa 7:20
VENTING 7:20
AIR DRYING 8:10
CYCLE COMPLETE 21:47

Digital Signature #
D74A0DD07A5A4EEB
