
Midmark M11 - v1.0.0
Total Cycles: 5022

Name: RITTER 1
Sterilizer ID: V2178198
10/24/2024 03:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 89.6 F 0.0
2:00 90.7 F 0.0
4:00 108.3 F 0.2
6:00 137.9 F 0.6
8:00 167.7 F 2.0
10:00 189.3 F 5.1
12:00 206.2 F 8.7
14:00 218.9 F 12.5
16:00 222.2 F 6.3
18:00 233.1 F 10.7
20:00 233.2 F 7.6
22:00 244.2 F 13.1
24:00 242.9 F 11.6
26:00 254.9 F 17.7
28:00 265.4 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.2 F 28.9
2:00 271.9 F 28.4
3:00 271.1 F 28.1
4:00 272.2 F 28.9

Min 271.1 F 28.0
Max 272.3 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.9 F 0.0
5:00 199.3 F 0.0
10:00 189.7 F 0.0
15:00 195.5 F 0.0
20:00 210.1 F 0.0
25:00 219.4 F 0.0
30:00 222.8 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 29:56
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:08:30
CYCLE COMPLETE

10/24/2024 04:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____