

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3945

Name: RITTER 1  
Sterilizer ID: V2178198  
10/25/2023 10:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 165.8 F 0.0  
2:00 163.6 F 2.3  
4:00 181.5 F 5.9  
6:00 200.3 F 11.1  
8:00 220.5 F 5.1  
10:00 231.3 F 10.1  
12:00 233.3 F 7.7  
14:00 243.7 F 12.6  
16:00 241.2 F 10.8  
18:00 254.2 F 17.3  
20:00 265.5 F 24.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.0  
1:00 272.7 F 29.1  
2:00 271.6 F 28.5  
3:00 272.4 F 28.5  
4:00 272.7 F 29.1  
  
Min 270.8 F 27.9  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.9 F 0.0  
5:00 246.0 F 0.0  
10:00 217.3 F 0.0  
15:00 226.4 F 0.0  
20:00 218.2 F 0.0  
25:00 215.6 F 0.0  
30:00 204.3 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 21:37  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:00:17  
CYCLE COMPLETE

10/25/2023 11:49 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_