
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000244
15:21 20/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:35
PRESSURIZING 3:36
STERILIZING 4:52
136.4 C 220kPa 4:52
136.1 C 223kPa 5:22
136.4 C 225kPa 5:52
136.4 C 224kPa 6:22
136.3 C 224kPa 6:52
136.3 C 223kPa 7:22
136.4 C 224kPa 7:52
136.3 C 224kPa 8:22
136.3 C 224kPa 8:22
VENTING 8:22
AIR DRYING 9:15
CYCLE COMPLETE 9:33

Digital Signature #
F21F94AADDBE8DD4
