
Midmark M11 - v1.0.5
Total Cycles: 2325

Name: RITTER2
Sterilizer ID: V2313228
05/02/2023 11:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 210.5 F 0.0
2:00 203.1 F 4.4
4:00 212.3 F 10.5
6:00 223.7 F 6.8
8:00 238.4 F 13.0
10:00 240.1 F 11.5
12:00 240.0 F 11.0
14:00 254.2 F 18.2
16:00 266.2 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.4 F 29.6
2:00 271.3 F 28.9
3:00 272.7 F 29.8
4:00 272.0 F 29.3

Min 271.2 F 28.8
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.1 F 0.0
5:00 215.5 F 0.0
10:00 232.8 F 0.0
15:00 230.7 F 0.0
20:00 223.1 F 0.0
25:00 225.2 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 2:28
HEATING: 17:34
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:57:19
CYCLE COMPLETE

05/02/2023 11:58 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____