

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1952

Name: RITTER 1  
Sterilizer ID: V2178198  
05/25/2022 01:41 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.3 F 0.0  
2:00 175.5 F 4.1  
4:00 193.0 F 8.4  
6:00 218.6 F 13.5  
8:00 228.1 F 8.4  
10:00 239.0 F 11.4  
12:00 239.2 F 10.8  
14:00 236.5 F 8.7  
16:00 250.9 F 15.5  
18:00 262.8 F 22.4  
20:00 272.1 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.1  
1:00 272.5 F 29.1  
2:00 272.0 F 29.0  
3:00 271.4 F 28.6  
4:00 272.1 F 28.6

Min 270.6 F 28.0  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.0 F 0.0  
5:00 232.6 F 0.0  
10:00 226.0 F 0.0  
15:00 227.3 F 0.0  
20:00 230.2 F 0.0  
25:00 221.4 F 0.0  
30:00 222.6 F 0.0  
Drying Complete

FILLING: 1:19  
HEATING: 20:05  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:58:31  
CYCLE COMPLETE

05/25/2022 02:39 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_