

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002207  
11:29 04/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:00  
PRESSURIZING 3:03  
STERILIZING 4:09  
136.4 C 225kPa 4:09  
136.2 C 225kPa 4:39  
136.4 C 226kPa 5:09  
136.4 C 226kPa 5:39  
136.1 C 225kPa 6:09  
136.1 C 223kPa 6:39  
136.4 C 227kPa 7:09  
136.4 C 226kPa 7:39  
136.4 C 226kPa 7:39  
VENTING 7:39  
AIR DRYING 8:31  
CYCLE COMPLETE 9:11  
-----

Digital Signature #  
1CC249A3CE76ADF3  
-----