
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002609
8:01 26/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:57
STERILIZING 4:05
136.2 C 221kPa 4:05
136.3 C 223kPa 4:35
136.4 C 224kPa 5:05
136.4 C 222kPa 5:35
136.2 C 221kPa 6:05
136.3 C 221kPa 6:35
136.3 C 221kPa 7:05
136.4 C 224kPa 7:35
136.4 C 224kPa 7:35
VENTING 7:35
AIR DRYING 8:26
CYCLE COMPLETE 9:01

Digital Signature #
2553ADCC356AE5F6
