
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002451
13:51 14/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:22
PRESSURIZING 1:26
STERILIZING 2:23
136.4 C 222kPa 2:23
136.2 C 222kPa 2:53
136.3 C 225kPa 3:23
136.4 C 224kPa 3:53
136.4 C 222kPa 4:23
136.4 C 223kPa 4:53
136.1 C 222kPa 5:23
136.2 C 224kPa 5:53
136.2 C 224kPa 5:53
VENTING 5:54
AIR DRYING 6:47
CYCLE COMPLETE 6:53

Digital Signature #
325E22DB4F005708
