
Midmark M11 - v1.0.5
Total Cycles: 1536

Name: RITTER2
Sterilizer ID: V2313228
09/26/2022 02:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.3 F 0.0
2:00 179.3 F 3.6
4:00 186.5 F 8.1
6:00 208.5 F 11.5
8:00 227.0 F 9.3
10:00 231.4 F 9.0
12:00 241.7 F 13.2
14:00 242.4 F 12.4
16:00 255.4 F 19.2
18:00 267.0 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 272.4 F 30.0
2:00 271.6 F 29.3
3:00 271.9 F 30.1
4:00 272.4 F 29.9

Min 271.1 F 29.2
Max 272.5 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 232.7 F 0.0
10:00 237.4 F 0.0
15:00 228.5 F 0.0
20:00 222.4 F 0.0
25:00 224.7 F 0.0
30:00 224.5 F 0.0

Drying Complete

FILLING: 1:36
HEATING: 19:27
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:18
CYCLE COMPLETE

09/26/2022 03:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____