
Midmark M11 - v1.0.0
Total Cycles: 1344

Name: RITTER 1
Sterilizer ID: V2178198
11/09/2021 09:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 96.5 F 0.0
2:00 100.7 F 0.0
4:00 126.0 F 0.0
6:00 150.7 F 0.1
8:00 191.0 F 1.0
10:00 205.4 F 4.8
12:00 217.7 F 8.4
14:00 228.5 F 12.4
16:00 230.5 F 7.9
18:00 242.3 F 13.1
20:00 240.8 F 10.9
22:00 237.9 F 9.3
24:00 252.2 F 16.2
26:00 263.9 F 23.0
28:00 271.5 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.5 F 28.7
2:00 271.5 F 28.0
3:00 272.6 F 28.8
4:00 271.8 F 28.1

Min 271.4 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 202.9 F 0.0
10:00 215.8 F 0.0
15:00 217.0 F 0.0
20:00 216.8 F 0.0
25:00 218.7 F 0.0
30:00 218.7 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 28:02
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:06:18
CYCLE COMPLETE

11/09/2021 10:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____