

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
2.3uS / 1.4ppm  
CYCLE NUMBER 003678  
8:48 25/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:42  
PRESSURIZING 1:49  
STERILIZING 2:43  
136.0 C 220kPa 2:43  
136.4 C 221kPa 3:13  
136.4 C 223kPa 3:43  
136.4 C 223kPa 4:13  
136.4 C 223kPa 4:43  
136.4 C 221kPa 5:13  
136.4 C 217kPa 5:43  
136.4 C 223kPa 6:13  
136.4 C 223kPa 6:13  
VENTING 6:13  
AIR DRYING 7:01  
CYCLE COMPLETE 7:33

-----  
Digital Signature #  
4AFC3A158179034C  
-----