
Midmark M11 - v1.0.0
Total Cycles: 4246

Name: RITTER 1
Sterilizer ID: V2178198
01/30/2024 08:53 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.5 F 0.0
2:00 74.8 F 0.3
4:00 107.9 F 1.1
6:00 108.3 F 3.1
8:00 135.9 F 5.7
10:00 160.9 F 8.8
12:00 182.4 F 12.4
14:00 202.1 F 2.5
16:00 212.6 F 5.3
18:00 222.6 F 8.6
20:00 232.3 F 12.3
22:00 232.4 F 7.9
24:00 243.7 F 13.0
26:00 241.5 F 10.9
28:00 253.1 F 16.7
30:00 263.1 F 22.6
32:00 272.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.1 F 28.4
2:00 272.5 F 28.9
3:00 271.9 F 28.6

Min 271.4 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.3 F 0.0
5:00 202.1 F 0.0
10:00 193.5 F 0.0
15:00 194.4 F 0.0
20:00 194.9 F 0.0
25:00 193.5 F 0.0
30:00 195.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 32:23
STERILIZING: 3:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:09:45
CYCLE COMPLETE

01/30/2024 10:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____