
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001259
16:51 15/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:45
PRESSURIZING 3:49
STERILIZING 5:03
136.0 C 221kPa 5:03
136.3 C 223kPa 5:33
136.4 C 225kPa 6:03
136.4 C 225kPa 6:33
136.4 C 222kPa 7:03
136.2 C 224kPa 7:33
136.0 C 223kPa 8:03
136.3 C 224kPa 8:33
136.3 C 224kPa 8:33
VENTING 8:33
AIR DRYING 9:31
CYCLE COMPLETE 24:31

Digital Signature #
E968F32D0A41330E
