

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2822

Name: RITTER2  
Sterilizer ID: V2313228  
08/31/2023 10:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.0 F 0.0  
2:00 159.8 F 0.0  
4:00 178.2 F 1.1  
6:00 201.8 F 5.7  
8:00 217.1 F 10.4  
10:00 224.8 F 6.5  
12:00 236.2 F 11.3  
14:00 235.1 F 9.0  
16:00 239.4 F 10.1  
18:00 247.9 F 14.4  
20:00 260.3 F 21.3  
22:00 271.4 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 272.2 F 28.9  
2:00 272.3 F 29.4  
3:00 271.7 F 28.6  
4:00 272.7 F 29.3

Min 271.4 F 28.5  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.5 F 0.0  
5:00 202.4 F 0.0  
10:00 200.9 F 0.0  
15:00 203.3 F 0.0  
20:00 209.4 F 0.0  
25:00 218.7 F 0.0  
30:00 219.8 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 22:37  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:00:43  
CYCLE COMPLETE

08/31/2023 11:13 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_