

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2734

Name: RITTER 1  
Sterilizer ID: V2178198  
01/04/2023 12:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.7 F 0.0  
2:00 191.3 F 0.0  
4:00 209.9 F 4.3  
6:00 219.5 F 9.6  
8:00 231.2 F 8.1  
10:00 238.0 F 11.6  
12:00 237.6 F 9.6  
14:00 237.5 F 8.7  
16:00 250.4 F 15.5  
18:00 263.0 F 22.7  
20:00 271.6 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.1 F 28.6  
2:00 272.2 F 29.0  
3:00 271.7 F 28.4  
4:00 272.6 F 29.1  
  
Min 271.4 F 28.2  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.6 F 0.0  
5:00 233.6 F 0.0  
10:00 226.7 F 0.0  
15:00 229.6 F 0.0  
20:00 232.8 F 0.0  
25:00 232.5 F 0.0  
30:00 228.2 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 20:04  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:06  
CYCLE COMPLETE

01/04/2023 01:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_