
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000744
11:54 23/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:37
STERILIZING 3:38
136.4 C 233kPa 3:38
136.3 C 233kPa 4:08
136.1 C 231kPa 4:38
136.4 C 233kPa 5:08
136.4 C 233kPa 5:38
136.3 C 232kPa 6:08
136.2 C 231kPa 6:38
136.3 C 232kPa 7:08
136.3 C 232kPa 7:08
VENTING 7:08
AIR DRYING 8:01
CYCLE COMPLETE 8:04

Digital Signature #
522A9B3B5E2603A4
