
Midmark M11 - v1.0.5
Total Cycles: 2952

Name: RITTER2
Sterilizer ID: V2313228
09/29/2023 01:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.6 F 0.0
2:00 167.1 F 3.2
4:00 188.0 F 7.3
6:00 203.9 F 12.0
8:00 217.5 F 6.6
10:00 230.1 F 11.5
12:00 231.8 F 8.8
14:00 238.2 F 10.5
16:00 245.8 F 13.9
18:00 257.8 F 20.4
20:00 268.8 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.7 F 30.0
2:00 271.6 F 29.3
3:00 272.1 F 30.1
4:00 272.4 F 29.7

Min 271.3 F 29.2
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.8 F 0.0
5:00 215.7 F 0.0
10:00 221.9 F 0.0
15:00 222.2 F 0.0
20:00 220.3 F 0.0
25:00 218.4 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 21:11
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:00:15
CYCLE COMPLETE

09/29/2023 02:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____