
Midmark M11 - v1.0.5
Total Cycles: 1173

Name: RITTER2
Sterilizer ID: V2313228
06/08/2022 10:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.2 F 0.0
2:00 175.8 F 2.9
4:00 182.7 F 7.1
6:00 205.1 F 12.6
8:00 219.8 F 8.2
10:00 233.1 F 11.6
12:00 238.2 F 11.5
14:00 237.8 F 10.0
16:00 251.3 F 16.6
18:00 262.9 F 23.5
20:00 272.6 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.3
1:00 272.5 F 30.1
2:00 271.8 F 29.4
3:00 271.2 F 29.5
4:00 272.5 F 30.1

Min 271.1 F 29.1
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 233.1 F 0.0
10:00 225.8 F 0.0
15:00 220.0 F 0.0
20:00 219.8 F 0.0
25:00 223.2 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 20:19
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:35
CYCLE COMPLETE

06/08/2022 11:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____