
Midmark M11 - v1.0.0
Total Cycles: 3273

Name: RITTER 1
Sterilizer ID: V2178198
05/24/2023 08:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.3 F 0.0
2:00 178.9 F 0.0
4:00 184.6 F 1.6
6:00 206.8 F 6.3
8:00 220.0 F 11.4
10:00 226.7 F 6.8
12:00 239.4 F 12.2
14:00 237.7 F 9.9
16:00 236.9 F 8.6
18:00 249.7 F 15.1
20:00 261.8 F 21.8
22:00 272.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.6 F 29.0
2:00 271.8 F 28.2
3:00 271.9 F 28.7
4:00 272.5 F 28.8

Min 271.3 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 224.5 F 0.0
10:00 239.8 F 0.0
15:00 230.5 F 0.0
20:00 226.9 F 0.0
25:00 227.9 F 0.0
30:00 230.6 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 22:26
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:59
CYCLE COMPLETE

05/24/2023 09:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____