

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 002270  
13:04 15/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:34  
PRESSURIZING 2:39  
STERILIZING 3:50  
136.4 C 225kPa 3:50  
136.2 C 224kPa 4:20  
136.4 C 225kPa 4:50  
136.4 C 223kPa 5:20  
136.2 C 224kPa 5:50  
136.4 C 224kPa 6:20  
136.2 C 220kPa 6:50  
136.2 C 223kPa 7:20  
136.2 C 223kPa 7:20  
VENTING 7:20  
AIR DRYING 8:09  
CYCLE COMPLETE 28:25  
-----

Digital Signature #  
084059EE509CB463  
-----