
Midmark M11 - v1.0.0
Total Cycles: 1474

Name: RITTER 1
Sterilizer ID: V2178198
12/20/2021 10:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 212.8 F 0.0
2:00 209.9 F 3.9
4:00 218.9 F 9.8
6:00 225.4 F 5.9
8:00 237.6 F 12.4
10:00 240.5 F 10.3
12:00 237.3 F 8.9
14:00 253.2 F 16.5
16:00 265.9 F 24.1
18:00 272.9 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 271.8 F 28.3
2:00 272.6 F 28.6
3:00 272.7 F 28.9
4:00 271.6 F 28.5

Min 271.4 F 28.2
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.0 F 0.0
5:00 241.7 F 0.0
10:00 229.8 F 0.0
15:00 229.4 F 0.0
20:00 228.7 F 0.0
25:00 230.1 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 18:13
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 00:56:57
CYCLE COMPLETE

12/20/2021 11:35 AM
ACCEPTED: X
REJECTED:
Operator: U02