
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000102
7:11 27/02/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:56
STERILIZING 4:16
136.5 C 223kPa 4:16
136.3 C 223kPa 4:46
136.4 C 223kPa 5:16
136.4 C 224kPa 5:46
136.2 C 221kPa 6:16
136.3 C 222kPa 6:46
136.3 C 223kPa 7:16
136.4 C 222kPa 7:46
136.4 C 222kPa 7:46
VENTING 7:46
AIR DRYING 8:39
CYCLE COMPLETE 8:44

Digital Signature #
D3A30209229F8141
