
Midmark M11 - v1.0.0
Total Cycles: 3982

Name: RITTER 1
Sterilizer ID: V2178198
11/06/2023 10:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.2 F 0.0
2:00 184.9 F 1.9
4:00 184.1 F 5.5
6:00 202.3 F 9.7
8:00 218.0 F 7.6
10:00 224.6 F 7.7
12:00 236.4 F 12.2
14:00 236.9 F 9.0
16:00 241.5 F 10.6
18:00 246.2 F 13.0
20:00 257.5 F 19.0
22:00 267.7 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.0
1:00 272.7 F 29.0
2:00 271.5 F 28.3
3:00 272.3 F 28.8
4:00 271.7 F 28.7

Min 270.9 F 27.9
Max 273.0 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.3 F 0.0
5:00 226.8 F 0.0
10:00 205.6 F 0.0
15:00 201.4 F 0.0
20:00 199.6 F 0.0
25:00 202.4 F 0.0
30:00 236.7 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 23:29
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:02:42
CYCLE COMPLETE

11/06/2023 11:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____