
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002111
14:41 02/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:02
STERILIZING 3:04
136.0 C 222kPa 3:04
136.4 C 224kPa 3:34
136.3 C 224kPa 4:04
136.2 C 224kPa 4:34
136.3 C 224kPa 5:04
136.4 C 223kPa 5:34
136.3 C 223kPa 6:04
136.2 C 223kPa 6:34
136.2 C 223kPa 6:34
VENTING 6:34
AIR DRYING 7:27
CYCLE COMPLETE 37:27

Digital Signature #
B6B209E03DFAC13F
