
Midmark M11 - v1.0.0
Total Cycles: 856

Name: RITTER 1
Sterilizer ID: V2178198
06/01/2021 06:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.6 F 0.0
2:00 78.3 F 0.0
4:00 91.5 F 0.0
6:00 119.6 F 0.0
8:00 148.9 F 0.1
10:00 169.3 F 0.1
12:00 189.7 F 0.2
14:00 208.2 F 3.0
16:00 214.0 F 5.7
18:00 222.2 F 8.7
20:00 231.2 F 12.1
22:00 231.0 F 7.5
24:00 241.9 F 12.1
26:00 238.7 F 9.8
28:00 239.8 F 9.7
30:00 248.3 F 14.0
32:00 259.2 F 20.0
34:00 269.0 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 27.9
1:00 271.9 F 28.3
2:00 272.7 F 28.6
3:00 272.0 F 28.0
4:00 271.6 F 27.9

Min 271.5 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.2 F 0.0
5:00 200.0 F 0.0
10:00 195.8 F 0.0
15:00 198.0 F 0.0
20:00 194.1 F 0.0
25:00 192.6 F 0.0
30:00 212.2 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 35:04
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:13:37
CYCLE COMPLETE

06/01/2021 07:58 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____