
Midmark M11 - v1.0.0
Total Cycles: 3760

Name: RITTER 1
Sterilizer ID: V2178198
09/12/2023 01:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.8 F 0.0
2:00 170.1 F 2.2
4:00 186.4 F 5.8
6:00 203.9 F 10.0
8:00 224.0 F 6.4
10:00 228.2 F 9.6
12:00 235.6 F 9.7
14:00 242.1 F 11.8
16:00 239.9 F 10.2
18:00 252.5 F 16.4
20:00 263.3 F 22.8
22:00 272.4 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.1 F 28.6
2:00 272.5 F 28.9
3:00 272.3 F 29.2
4:00 271.3 F 28.6

Min 271.3 F 28.5
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.1 F 0.0
5:00 239.6 F 0.0
10:00 225.0 F 0.0
15:00 203.3 F 0.0
20:00 227.6 F 0.0
25:00 232.4 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 22:54
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:02:06
CYCLE COMPLETE

09/12/2023 02:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____